Ref. Certif. No.

DK-77173-M1-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Multichannel Data Acquisition Module
Name and address of the applicant	Best Research & Development Limited UNIT 3-4, ST AUSTELL ENTERPRISE PARK TREVERBYN RD ST AUSTELL, PL25 4EJ United Kingdom
Name and address of the manufacturer	Best Research & Development Limited UNIT 3-4, ST AUSTELL ENTERPRISE PARK TREVERBYN RD ST AUSTELL, PL25 4EJ United Kingdom
Name and address of the factory Note: *When more than one factory, please report on page 2*	BLUECHIPS MICROHOUSE CO LTD 155 MOO 5 T CHAE CHANG, A SANKAMPHAENG CHIANG MAI, 50130 Thailand ☐ Additional Information on page 2
Ratings and principal characteristics	100-240 V ac, 50/60 Hz, 20 W Max; Measurement 3-phase 346 L-N V rms, 600 L-L V rms, Cat III
Trademark (if any)	None
Type of Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	Eniscope 0 Hybrid, Eniscope 4 Hybrid, Eniscope 8 Hybrid, Eniscope Q Hybrid 0, Eniscope Q Hybrid 4, Eniscope Q Hybrid 8
Additional information (if necessary may also be reported on page 2)	The report was revised to include technical modifications. ☒ Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010, IEC 61010-2-030:2010
As shown in the Test Report Ref. No. which forms part of this Certificate	E464761-D1000-1/A1/C0 issued on 2019-06-06

This CB Test Certificate is issued by the National Certification Body

☐ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
☒ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2019-06-12
Original Issue Date: 2018-10-29

Signature:

Jan-Erik Storgaard

Additional Information:
The original report was modified to include the following changes/additions:
- Updated model designations, enclosure drawings, marking labels, Test item description and electrical ratings and photographs.
- Updated CCT with corrected schematic designators for capacitors, and incorrect part numbers.
- Updated test plan to include all applicable tests from original evaluation, including waived tests.

All updates are only corrections and no construction changes.

National Differences specified in the CB Test Report.

Additional information (if necessary)

☐ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
☒ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2019-06-12
Original Issue Date: 2018-10-29

Signature:
Jan-Erik Storgaard